

SLUŽBENA OZNAKA TIP A	M-1-6
-----------------------	-------

Š I F R A	
-----------	--

VRSTA MJERILA	Vaga „analitička“ elektronska																														
RAZ. TOČNOSTI	①																														
MJERNI OPSEG	20g; 5g/20g; 10g; 6g/20g																														
PROIZVOĐAČ	„Mettler“ Zürich, Švicarska																														
ODOBRENI TIPOVI	<table border="1"> <thead> <tr> <th></th> <th>max</th> <th>Min</th> <th>e</th> <th>d</th> </tr> </thead> <tbody> <tr> <td>AT 201</td> <td>20g</td> <td>1mg</td> <td>1mg</td> <td>0,1mg</td> </tr> <tr> <td>AT 250</td> <td>5g/20g</td> <td>10mg</td> <td>1mg</td> <td>0,01/0,1mg</td> </tr> <tr> <td>AT 200</td> <td>20g</td> <td>10mg</td> <td>1mg</td> <td>0,1mg</td> </tr> <tr> <td>AT 100</td> <td>10g</td> <td>10mg</td> <td>1mg</td> <td>0,1mg</td> </tr> <tr> <td>AT 261</td> <td>6g/20g</td> <td>10mg</td> <td>1mg</td> <td>0,01/0,1mg</td> </tr> </tbody> </table> <p>i ostale AT ...</p>		max	Min	e	d	AT 201	20g	1mg	1mg	0,1mg	AT 250	5g/20g	10mg	1mg	0,01/0,1mg	AT 200	20g	10mg	1mg	0,1mg	AT 100	10g	10mg	1mg	0,1mg	AT 261	6g/20g	10mg	1mg	0,01/0,1mg
	max	Min	e	d																											
AT 201	20g	1mg	1mg	0,1mg																											
AT 250	5g/20g	10mg	1mg	0,01/0,1mg																											
AT 200	20g	10mg	1mg	0,1mg																											
AT 100	10g	10mg	1mg	0,1mg																											
AT 261	6g/20g	10mg	1mg	0,01/0,1mg																											
GLASNIK br/god/str	4/89/40																														
PRI ISPITIVANJU TIP A MJERIL A ZADOVOLJEL A UVJETE PROPISANE PRAVILNIKOM ZA _____																															
Odobrenje postoji u arhivi KM Zagreb	DA - <input checked="" type="radio"/> NE																														